

Supporting Information

J.Chem. Metrol. 11:2 (2017) 61-67

Development and validation of a simple HPLC assay method for ibuprofen, pseudoephedrine hydrochloride and chlorpheniramine maleate in syrup form

Hande Kazak Sarılmışer* , N. Orçun Özçelik, Banu Özkırım Arslan, Mine Gökalp, Arzu Ustaoğlu, Udaya Kumar Dude and Zdravka Knezevic

Abdi Ibrahim Pharmaceutical Company, Research and Development Center, Esenyurt, İstanbul, Türkiye

Table of Contents	Page
Table S1. Information of standard compounds	2
Table S2. Gradient program for chromatographic conditions.	2
Table S3. Linearity and range – IBU (1)	3
Figure S1. Linearity of IBU (1)	3
Table S4. Linearity and range – PSEH (2)	4
Figure S2. Linearity of PSEH (2)	4
Table S5. Linearity and range – CHL (3)	5
Figure S3. Linearity of CHL (3)	5
Figure S4. HPLC chromatogram of diluent	6
Figure S5. HPLC chromatogram of standards	7
Table S6. Repeatability (precision of the method) results of IBU, PSEH and CHL	9
Table S7. Intermediate precision results of IBU, PSEH and CHL	9
Table S8. Comparison of repeatability and intermediate precision results of IBU, PSEH and CHL	10
Table S9. Comparison of robustness and repeatability test results – IBU (1)	11
Table S10. Comparison of robustness and repeatability test results – PSEH (2)	11
Table S11. Comparison of robustness and repeatability test results – CHL (3)	11

Table S1. Information of standard compounds

Standard Name	Lot Number	Source	Potency% (as is)
IBU (1)	C100-1004161M	Hubei Biocause Pharmaceutical Co., Ltd. (China)	99.4
PSEH (2)	510111Z02	Malladi Drugs & Pharmaceuticals Ltd. (India)	99.9
CHL (3)	SLL/C/0911150	Supriya Lifescience Ltd. (India)	99.5

Table S2. Gradient program for chromatographic conditions

Time (min.)	Mobile Phase A (%) (%)	Mobile Phase B (%)
0-7	50	50
7-13	50→40	50→60
13-17	40	60
17-26	40→30	60→70
26-28	30→50	70→50
28-30	50	50

Table S3. Linearity and range – IBU (1)

Linearity Level (%)	Concentration (mg/mL)	Peak Area	Average Area
70	0.209997	7734457	7726879
		7719301	
80	0.239997	8820499	8828219
		8835939	
90	0.269996	9919461	9918685
		9917908	
100	0.299996	11104898	11108852
		11112805	
120	0.359995	13178299	13182496
		13186693	
130	0.389994	14281307	14298248
		14315188	


**Figure S1.** Linearity of IBU (1)

Table S4. Linearity and range – PSEH (2)

Linearity Level (%)	Concentration (mg/mL)	Peak Area	Average Area
70	0.031503	831763	830730
		829697	
80	0.036004	948172	948157
		948142	
90	0.040504	1067238	1067508
		1067777	
100	0.045005	1188584	1188896
		1189208	
120	0.054006	1415834	1416092
		1416349	
130	0.058506	1526714	1529258
		1531801	


**Figure S2.** Linearity of PSEH (2)

Table S5. Linearity and range – CHL (3)

Linearity Level (%)	Concentration (mg/mL)	Peak Area	Average Area
70	0.002102	51201	
		51572	51387
80	0.002402	58997	
		58730	58864
90	0.002703	66172	
		66107	66140
100	0.003003	74076	
		74089	74083
120	0.003603	88351	
		88470	88411
130	0.003904	96571	
		96780	96676




Figure S3. Linearity of CHL (3)




Figure S4. HPLC chromatogram of diluent




Figure S5. HPLC chromatogram of standards

Table S6. Repeatability (precision of the method) results of IBU, PSEH and CHL

Sample No	IBU (1) Assay %	PSEH (2) Assay %	CHL (3) Assay %
1	98.0	97.7	97.8
2	97.5	97.6	97.5
3	97.6	97.6	97.7
4	97.2	98.0	98.0
5	98.2	97.5	97.5
6	98.1	98.0	97.8
Average	97.8	97.7	97.7
SD	0.393	0.216	0.194
RSD%	0.41	0.22	0.20
Confidence Interval (CI 95%)	97.8 ± 0.30	97.7± 0.20	97.7 ± 0.20

Table S7. Intermediate precision results of IBU, PSEH and CHL

Sample No	IBU (1) Assay %	PSEH (2) Assay %	CHL (3) Assay %
1	97.3	98.3	98.2
2	98.2	97.9	97.9
3	97.7	97.5	97.6
4	97.9	97.7	97.8
5	97.8	97.5	97.5
6	97.1	97.9	97.6
Average	97.7	97.8	97.8
SD	0.40	0.30	0.26
RSD%	0.41	0.31	0.26
Confidence Interval (CI 95%)	97.7 ± 0.30	97.8 ± 0.20	97.8 ± 0.20

Table S8. Comparison of repeatability and intermediate precision results of IBU, PSEH and CHL

	Sample No	IBU (1) Assay %	PSEH (2) Assay %	CHL(3) Assay %
Repeatability (precision of the method)	1	98.0	97.7	97.8
	2	97.5	97.6	97.5
	3	97.6	97.6	97.7
	4	97.2	98.0	98.0
	5	98.2	97.5	97.5
	6	98.1	98.0	97.8
Intermediate precision	1	97.3	98.3	98.2
	2	98.2	97.9	97.9
	3	97.7	97.5	97.6
	4	97.9	97.7	97.8
	5	97.8	97.5	97.5
	6	97.1	97.9	97.6
Average	97.8	97.7	97.7	
SD	0.25	0.22	0.38	
RSD%	0.26	0.22	0.39	
Confidence Interval (CI 95%)	97.8 ± 0.10		97.7 ± 0.10	
			97.7 ± 0.20	

Table S9. Comparison of robustness and repeatability test results – IBU (1)

Name of Analysis	Assay %	Retention Time (min)	Symmetry Factor
Repeatability	97.8	25.816	1.1
Flow rate: 1.0 mL/min	98.1	27.956	1.1
Flow rate: 1.4 mL/min	98.4	24.595	1.2
Wavelength: 212 nm	97.3	25.679	1.1
Wavelength: 216 nm	97.2	25.671	1.1
Column Temperature: 48°C	98.4	26.240	1.3
Column Temperature: 52°C	98.8	25.334	1.3

Table S10. Comparison of robustness and repeatability test results – PSEH (2)

Name of Analysis	Assay %	Retention Time (min)	Symmetry Factor
Repeatability	97.7	8.997	0.9
Flow rate: 1.0 mL/min	97.3	10.965	0.9
Flow rate: 1.4 mL/min	97.3	7.966	1.1
Wavelength: 212 nm	97.3	9.151	0.9
Wavelength: 216 nm	97.0	9.142	0.9
Column Temperature: 48°C	98.0	9.193	1.2
Column Temperature: 52°C	98.3	8.372	1.1

Table S11. Comparison of robustness and repeatability test results – CHL (3)

Name of Analysis	Assay %	Retention Time (min)	Symmetry Factor
Repeatability	97.7	19.980	1.0
Flow rate: 1.0 mL/min	98.4	22.098	1.0
Flow rate: 1.4 mL/min	97.5	18.287	1.1
Wavelength: 212 nm	97.8	20.070	1.0
Wavelength: 216 nm	98.1	20.061	1.0
Column Temperature: 48°C	98.3	19.856	1.1
Column Temperature: 52°C	98.4	18.852	1.2