

Supporting Information

J. Chem. Metrol. 17:2 (2023) 181-198

Establishment of HPLC method for evaluation of process related impurities of letermovir and LCMS/MS characterization of forced degradation compounds

**Rajesh Varma Bhupatiraju^{1*}, B Srinivasa Kumar², Pavani Peddi³
and Venkata Swamy Tangeti⁴**


¹*Department of Chemistry, GITAM Institute of Science, GITAM (Deemed to be University),
Visakhapatnam – 530 045, India*

²*Associate Professor, Department of Mathematics, Koneru Lakshmaiah Education Foundation,
Vaddeswaram, Guntur (D.T), A.P., 522302*

³*Department of Chemistry, P.V.P. Siddhartha Institute of Technology, Vijayawada-520007, India*


⁴*Assistant Professor, Department of Chemistry, Tagore Government Arts and Science College
(Affiliated to Pondicherry University), Puducherry*

Table of Contents	Page
Figure S1: Details of impurities in the study	2
Figure S2A: Mass spectrum of DP 1	3
Figure S2B: Mass spectrum of DP 2	3
Figure 2C: Mass spectrum of DP 3	4
Figure S2D: Mass spectrum of DP 4	4
Figure S2E: Mass spectrum of DP 5	5
Table S1: Uncertainty assessment of letermovir	6
Table S2: Uncertainty assessment of Impurity 1	7
Table 3: Uncertainty assessment of Impurity 2	8
Table S4: Uncertainty assessment of Impurity 3	9
Table S5: Uncertainty assessment of Impurity 4	10




Impurity 1

Name: *(S)*-2-(3-(2-Methoxy-5-(trifluoromethyl)phenyl)-2,8-bis(4-(3-methoxyphenyl)piperazin-1-yl)-3,4-dihydroquinazolin-4-yl)acetic acid;
Formula: C₄₀H₄₃F₃N₆O₅; **Mass:** 744.8 g/mol




Impurity 2

Name: *(S)*-2-(8-Fluoro-3-(2-hydroxy-5-(trifluoromethyl)phenyl)-2-(4-(3-methoxyphenyl)piperazin-1-yl)-3,4-dihydroquinazolin-4-yl)acetic acid;
Formula: C₂₈H₂₆F₄N₄O₄; **Mass:** 558.5 g/mol



Impurity 3

Name: *(S)*-2-(8-Fluoro-2-(4-(3-hydroxyphenyl)piperazin-1-yl)-3-(2-methoxy-5-(trifluoromethyl)phenyl)-3,4-dihydroquinazolin-4-yl)acetic acid;
Formula: C₂₈H₂₆F₄N₄O₄; **Mass:** 558.5 g/mol



Impurity 4

Name: *(S)*-2-(8-Fluoro-3-(2-methoxy-5-(trifluoromethyl)phenyl)-2-(piperazin-1-yl)-3,4-dihydroquinazolin-4-yl)acetic acid;
Formula: C₂₂H₂₂F₄N₄O₃; **Mass:** 466.4 g/mol

Figure S1: Details of impurities in the study




Figure S2A: Mass spectrum of DP 1




Figure S2B: Mass spectrum of DP 2




Figure S2C: Mass spectrum of DP 3




Figure S2D: Mass spectrum of DP 4




Figure S2E: Mass spectrum of DP 5

Table S1: Uncertainty Assessment of Ietermovir

Uncertainty in standard sample		X	Xi	N	U_{standard} Formula	U_{standard}
		100	98.75	3	$\frac{X - Xi}{\sqrt{n}}$	0.722
Uncertainty in Slope of Calibration Plot		SD	n	Standard Error	$U_{\text{Calibration}}$ Formula	$U_{\text{Calibration}}$
Slope 1	9837.3	7.65	3	4.419	$\frac{\text{Standard Error of Slope} \times 100}{\text{Slope}}$	0.045
Slope 2	9826.8					
Slope 3	9841.7					
Uncertainty in Recovery of Sample		SD	n	Standard Error	U_{Recovery} Formula	U_{Recovery}
Recovery 1	98.63	0.597	6	0.244	$\frac{\text{Standard Error of Recovery} \times 100}{\text{Recovery}}$	0.247
Recovery 2	98.14					
Recovery 3	99.47					
Recovery 4	98.59					
Recovery 5	99.03					
Recovery 6	99.74					
Uncertainty associated with repeatability		SD	n	Standard Error	$U_{\text{Repeatability}}$ Formula	$U_{\text{Repeatability}}$
Repeatability 1	149.58	0.496	6	0.202	$\frac{\text{Standard Error of Repeatability} \times 100}{\text{Repeatability}}$	0.136
Repeatability 2	149.63					
Repeatability 3	148.99					
Repeatability 4	150.52					
Repeatability 5	149.57					
Repeatability 6	149.85					
					U_{Combined} Formula	U_{Combined}
					$\sqrt{(U_{\text{standard}})^2 + (U_{\text{Calibration}})^2 + (U_{\text{Recovery}})^2 + (U_{\text{Repeatability}})^2}$	0.776
					U_{Expanded} Formula	U_{Expanded}
					$U_{\text{Combined}} \times 1.96$	1.521

Table S2: Uncertainty Assessment of Impurity 1

Uncertainty in standard sample		X	Xi	N	U_{standard} Formula	U_{standard}
		100	99.32	3	$\frac{X - Xi}{\sqrt{n}}$	0.393
Uncertainty in Slope of Calibration Plot		SD	n	Standard Error	$U_{\text{Calibration}}$ Formula	$U_{\text{Calibration}}$
Slope 1	801787	131.50	3	75.93	$\frac{\text{Standard Error of Slope} \times 100}{\text{Slope}}$	0.009
Slope 2	801524					
Slope 3	801658					
Uncertainty in Recovery of Sample		SD	n	Standard Error	U_{Recovery} Formula	U_{Recovery}
Recovery 1	98.74	1.186	6	0.484	$\frac{\text{Standard Error of Recovery} \times 100}{\text{Recovery}}$	0.488
Recovery 2	98.34					
Recovery 3	98.45					
Recovery 4	100.58					
Recovery 5	100.95					
Recovery 6	98.44					
Uncertainty associated with repeatability		SD	n	Standard Error	$U_{\text{Repeatability}}$ Formula	$U_{\text{Repeatability}}$
Repeatability 1	0.1498	0.0019	6	0.00078	$\frac{\text{Standard Error of Repeatability} \times 100}{\text{Repeatability}}$	0.531
Repeatability 2	0.1463					
Repeatability 3	0.1488					
Repeatability 4	0.1479					
Repeatability 5	0.1465					
Repeatability 6	0.1445					
					U_{Combined} Formula	U_{Combined}
					$\sqrt{(U_{\text{standard}})^2 + (U_{\text{Calibration}})^2 + (U_{\text{Recovery}})^2 + (U_{\text{Repeatability}})^2}$	0.821
					U_{Expanded} Formula	U_{Expanded}
					$U_{\text{Combined}} \times 1.96$	1.609

Table S3: Uncertainty Assessment of Impurity 2

Uncertainty in standard sample		X	Xi	N	U_{standard} Formula	U_{standard}				
		100	98.75	3	$\frac{X - Xi}{\sqrt{n}}$	0.722				
Uncertainty in Slope of Calibration Plot		SD	n	Standard Error	$U_{\text{Calibration}}$ Formula	$U_{\text{Calibration}}$				
Slope 1	370263	416.34	3	240.37	$\frac{\text{Standard Error of Slope} \times 100}{\text{Slope}}$	0.065				
Slope 2	370114									
Slope 3	370898									
Uncertainty in Recovery of Sample		SD	n	Standard Error	U_{Recovery} Formula	U_{Recovery}				
Recovery 1	99.36	0.494	6	0.201	$\frac{\text{Standard Error of Recovery} \times 100}{\text{Recovery}}$	0.202				
Recovery 2	100.41									
Recovery 3	100.32									
Recovery 4	99.75									
Recovery 5	99.82									
Recovery 6	99.18									
Uncertainty associated with repeatability		SD	n	Standard Error	$U_{\text{Repeatability}}$ Formula	$U_{\text{Repeatability}}$				
Repeatability 1	0.1485	0.00076	6	0.00031	$\frac{\text{Standard Error of Repeatability} \times 100}{\text{Repeatability}}$	0.209				
Repeatability 2	0.1479									
Repeatability 3	0.1474									
Repeatability 4	0.1489									
Repeatability 5	0.1494									
Repeatability 6	0.1491									
					U_{Combined} Formula	U_{Combined}				
					$\sqrt{(U_{\text{standard}})^2 + (U_{\text{Calibration}})^2 + (U_{\text{Recovery}})^2 + (U_{\text{Repeatability}})^2}$					0.781
					U_{Expanded} Formula					U_{Expanded}
					$U_{\text{Combined}} \times 1.96$					1.530

Table S4: Uncertainty Assessment of Impurity 3

Uncertainty in standard sample		X	Xi	N	U_{standard} Formula	U_{standard}
		100	99.25	3	$\frac{X - Xi}{\sqrt{n}}$	0.433
Uncertainty in Slope of Calibration Plot		SD	n	Standard Error	$U_{\text{Calibration}}$ Formula	$U_{\text{Calibration}}$
Slope 1	664903	324.84	3	187.55	$\frac{\text{Standard Error of Slope} \times 100}{\text{Slope}}$	0.028
Slope 2	664258					
Slope 3	664513					
Uncertainty in Recovery of Sample		SD	n	Standard Error	U_{Recovery} Formula	U_{Recovery}
Recovery 1	99.63	0.367	6	0.150	$\frac{\text{Standard Error of Recovery} \times 100}{\text{Recovery}}$	0.151
Recovery 2	99.74					
Recovery 3	99.37					
Recovery 4	98.85					
Recovery 5	98.91					
Recovery 6	99.43					
Uncertainty associated with repeatability		SD	n	Standard Error	$U_{\text{Repeatability}}$ Formula	$U_{\text{Repeatability}}$
Repeatability 1	0.1496	0.00082	6	0.00033	$\frac{\text{Standard Error of Repeatability} \times 100}{\text{Repeatability}}$	0.224
Repeatability 2	0.1491					
Repeatability 3	0.1498					
Repeatability 4	0.1486					
Repeatability 5	0.1503					
Repeatability 6	0.1509					
					U_{Combined} Formula	U_{Combined}
					$\sqrt{(U_{\text{standard}})^2 + (U_{\text{Calibration}})^2 + (U_{\text{Recovery}})^2 + (U_{\text{Repeatability}})^2}$	0.511
					U_{Expanded} Formula	U_{Expanded}
					$U_{\text{Combined}} \times 1.96$	1.002

Table S5: Uncertainty Assessment of Impurity 4

Uncertainty in standard sample		X	Xi	N	U_{standard} Formula	U_{standard}
		100	99.65	3	$\frac{X - Xi}{\sqrt{n}}$	0.202
Uncertainty in Slope of Calibration Plot		SD	n	Standard Error	$U_{\text{Calibration}}$ Formula	$U_{\text{Calibration}}$
Slope 1	474396	1626.67	3	939.16	$\frac{\text{Standard Error of Slope} \times 100}{\text{Slope}}$	0.198
Slope 2	474029					
Slope 3	471413					
Uncertainty in Recovery of Sample		SD	n	Standard Error	U_{Recovery} Formula	U_{Recovery}
Recovery 1	99.44	0.397	6	0.162	$\frac{\text{Standard Error of Recovery} \times 100}{\text{Recovery}}$	0.162
Recovery 2	100.59					
Recovery 3	100.12					
Recovery 4	100.24					
Recovery 5	99.88					
Recovery 6	99.79					
Uncertainty associated with repeatability		SD	n	Standard Error	$U_{\text{Repeatability}}$ Formula	$U_{\text{Repeatability}}$
Repeatability 1	0.1487	0.00139	6	0.00057	$\frac{\text{Standard Error of Repeatability} \times 100}{\text{Repeatability}}$	0.381
Repeatability 2	0.1493					
Repeatability 3	0.1476					
Repeatability 4	0.1509					
Repeatability 5	0.1512					
Repeatability 6	0.1504					
					U_{Combined} Formula	U_{Combined}
					$\sqrt{(U_{\text{standard}})^2 + (U_{\text{Calibration}})^2 + (U_{\text{Recovery}})^2 + (U_{\text{Repeatability}})^2}$	0.502
					U_{Expanded} Formula	U_{Expanded}
					$U_{\text{Combined}} \times 1.96$	0.983

